

COVID-19 ACIST RXi® and RXi Mini™ Cleaning

April 1, 2020

Dear Valued Customer,

In anticipation of questions arising around COVID-19 and ACIST RXi® and RXi Mini™, ACIST Medical would like to provide the following:

- **RXi Cleaning Procedure:** The RXi User Guide and the RXi Mini User Guide provide detailed instructions on how best to clean their respective systems. Please review the User Guides at <http://acist.com/library>.
 - **Username:** acist
 - **Password:** thepowerto
 - **Select the Product:** “RXi Rapid Exchange System (USA)”, “RXi Rapid Exchange FFR System (International)”, or “ACIST RXi Mini” and then your preferred language

Although specific testing for COVID-19 has not been executed, please note these key recommendations. Do not use detergents, cleaning solutions, isopropyl alcohol, steam, any oils or materials containing oil. The recommendation is to use a 10% bleach solution. For a detailed, step-by-step process please consult pages 76-77 of the US RXi User Guide, 77-78 of the International RXi User Guide, or 63-64 of the RXi Mini User Guide.

We at ACIST Medical would like to thank you for your efforts during this time of uncertainty. Without the efforts of amazing healthcare and hospital staff such as you and your teams, access to superb patient care would not be possible during this unprecedented time.

Please know, that in this time of great uncertainty, we are committed to supporting your pursuit of patient care. Should you have additional questions or need further assistance, please contact your ACIST sales representative or Customer Support at +1-888-670-7701 for the US or +31 (0)45-750 700/dial1 for Europe.

Best regards,



Kristen Knox
Sr. Vice President, General Manager
ACIST Medical Systems, Inc.

PN: 30-100074